



# New City Scales Ltd

**Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service**

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,  
Milton Keynes, Bucks MK17 0PD  
Tel: 01296 714477  
Fax: 01296 714483  
E-mail: support@newcityscales.co.uk  
Web Site: www.newcityscales.co.uk

CERTIFICATE SERIAL NO: 9690

DATE: 21<sup>st</sup> March 2020

## QUALITY ASSURANCE

### CERTIFICATE OF TEST

CUSTOMER: Sharp Skips Ltd

YOUR PURCHASE ORDER:

ADDRESS: Ferry Lane  
Rainham  
Essex  
RM13 9HA


TYPE:  
Vishay VT300

SERIAL NO  
T16310304

CAPACITY:  
50,000kg

We certify that the above instrument/equipment has been subjected to our full test procedures, and conforms to the original manufacturer's specification and/or appropriate British, National or International Standard; and is in accordance with your stated order requirements.

We further certify that the above instrument/equipment has been tested using equipment and standards which are traceable to National Standards. By virtue of the Weights and Measures Act 1985.

SIGNED:   
For and on Behalf of New City Scales.

POSITION: 

DATE: 21-3-20





# New City Scales Ltd

**Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service**

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,  
 Milton Keynes, Bucks MK17 0PD  
 Tel: 01296 714477  
 Fax: 01296 714483  
 E-mail: support@newcityscales.co.uk  
 Web Site: www.newcityscales.co.uk

## CALIBRATION RECORD

**TYPE:** VT300

**MANUFACTURE:** Vishay

**SERIAL:** T16310304

**CAPACITY / INCREMENTS:** 50.000kg x 20kg

**ALLOWABLE TOLERANCE** + - 20kg

### TEST READINGS

CALIBRATION WEIGHT	READING	ERROR
ZERO	0	_____
2,000kg	2,000kg	_____
6,000kg	6,000kg	_____
10,000kg	10,000kg	_____
12,000kg	12,000kg	_____
16,000kg	16,000kg	_____
20,000kg	20,000kg	_____
20,400kg	20,400kg	_____
Substituted with dead-weight	20,100kg	_____
30,100kg	30,100kg	_____
40,100kg	40,100kg	_____
44,560kg	44.560kg	_____

Individual load-cell test. No 1 = 16,000kg. 2 = 16,000kg. 3 = 16,000kg. 4 = 16,000kg.

**CALIBRATION WEIGHTS USED:** Supplied by New City Scales Cert no: TR10945 / TR10946

**TESTED BY NEW CITY SCALES LTD**

**SIGNED**.....

**DATED : 21<sup>st</sup> March 2020**

